

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI) – Italia – Strada dei Laghi 39
Tel +39-0577-319554 Fax +39-0577-319020



Sede Legale e Amministrazione
20144 Milano (Italy) Via A. Solari 19 sc. 6

QUALITY CONTROL CERTIFICATE

KIT NAME: 88000-C286 - IBUPROFEN

Code	88000 - C286			
Lot No.	930	2024-04	2023-06	

REAGENTS

Label	Description	Lot.	Exp	No
D5841	88000- C286 IBUPROFEN	930	2024-04	2X6
B5840	CALIBRATORE 88000- C286 IBUPROFEN	930	2024-06	1
C5840	CONTROL SERUM 88000- C286 IBUPROFEN	930	2024-06	1

CONTROLS

Found results	Expected results
---------------	------------------


Calibrator (mOD)	279	≥ 100
Control Serum (KU/L)	17,5	10,1-44,7

CALIBRATOR

7	2	1	9	3	0	6	5	2	0	6	2	1	8	7	4	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES

Operator Signature *RB*

CONTROL SERUM

7	2	1	9	3	0	6	3	9	7	2	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES

Operator Signature *RB*

The test has been carried out in accordance with the package insert

TESTED BY: Simone Baiocchi

APPROVED

RESPONSIBLE HEAD, Quality Control Reagents.....

DATE 06/06/2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested by an FDA approved method and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.