DIESSE DIAGNOSTICA SENESE SpA

53035 Monteriggioni (SI), Italy, Strada dei Laghi 39 Tel. +39 0577319554 FAX +39 0577319020




Sede Legale 20144 Milano (Italy) Via A. Solari 19 sc. 6

Code 86050 Lot No. 928 2024-04 QUALITY CONTROL CERTIFICATE M 2023-04

CONTROLS				
C3351	CONTROL SERUM BETA 2-GLYCOPROTEIN-G	928	2024-04	1
B3351	CALIBRATORE BETA 2-GLYCOPROTEIN-G	928	2024-04	1
D3351	BETA 2-GLYCOPROTEIN-G	928	2024-04	6X6

Foun	d results	Expected results
------	-----------	------------------

Calibrator (mOD)	497	≥ 100
Control Serum (AU/mL)	33,3	23,1-50,0



CALIBRATOR

2 0 6 9 2 8 6 1 2 3 3 3 5 6 5 4 0 0

Label checked with barcode reader

YES 🔀

Operator Signature



2 0 6 9 2 8 6 1 3 2 3 0 0 0 0 0

Label checked with barcode reader

YES 🔀

0



This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents...

DATE 66 06/23

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.