

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME HERPES SIMPLEX 1+2 IgG -CHORUS

Code 81020

Lot No. 936



2024-08



2023-05

REAGENTS

Label	Description	Lot.	Exp.	NO
D2211	Herpes simplex 1+2 IgG	936	2024-08	6X6
B2212	CALIBRATORE Herpes simplex 1+2 IgG	936	2024-08	1
C2211	CONTROL SERUM Herpes simplex 1+2 IgG	936	2024-08	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	405	≥ 100
Control Serum (INDEX)	3,0	2,0-3,7
Negative Control (INDEX)	0,1	$<0,7$

CALIBRATOR


0	0	4	9	3	6	7	0	0	0	0	0	3	6	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature *RB*



CONTROL SERUM


0	0	4	9	3	6	7	2	0	3	7	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature *RB*



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents..... *Abuse*

DATE 09/06/2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.