

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME CHORUS DENGUE IgM CAPTURE

Code	81157		
Lot No.	934	2024-05	2023-05

REAGENTS

Label	Description	Lot.	Exp	No
D4140	Dengue IgM Capture	934	2024-05	6X6
B4140	Calibratore Dengue IgM Capture	934	2024-05	1
C4140	Control Dengue IgM Capture	934	2024-05	1

CONTROLS

Found results	Expected results
---------------	------------------

Calibrator (mOD)	312	> 100
Control (Index)	3,7	2,2-4,1

CALIBRATOR


9	8	6	9	3	4	6	0	0	0	0	0	3	2	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature



CONTROL SERUM


9	8	6	9	3	4	6	2	2	4	1	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature



This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents

DATE 08.06.2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.