

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME CHORUS PROMONITOR ANTI-USTEKINUMAB

Code 86713

Lot No. 934



2024-05



2023-05

REAGENTS

Label	Description	Lot.	Exp.	NO
D4510	PROMONITOR ANTI-USTEKINUMAB	934	2024-05	6x6
B4510	CALIBRATORE ANTI-USTEKINUMAB	934	2025-02	1
C4510	CONTROL SERUM ANTI-USTEKINUMAB	934	2025-02	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	260	≥ 100
Control Serum (AU/mL)	14,4	12,6-18,3

CALIBRATOR


2	8	1	9	3	4	6	7	2	0	4	3	3	1	7	5	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature *tb*



CONTROL SERUM

2	8	1	9	3	4	6	2	6	3	9	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature *tb*



This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents *Blue*

DATE 13.06.2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.