

DIESSE DIAGNOSTICA SENESE SpA  
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39  
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale  
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

## QUALITY CONTROL CERTIFICATE

KIT NAME CHORUS PHOSPHOLIPID SCREEN IgM

Code 86108

Lot No. 938



2024-05



2023-05

### REAGENTS

Label	Description	Lot.	Exp.	NO
D4020	Phospholipid Screen IgM	938	2024-05	6X6
B4020	CALIBRATORE Phospholipid Screen IgM	938	2024-05	1
C4020	CONTROL SERUM Phospholipid Screen IgM	938	2024-05	1

### CONTROLS


	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	390	$\geq 100$
Control (AU/ml)	38,3	26,0-83,6

### CALIBRATOR

9	7	5	9	3	8	6	1	1	9	2	9	3	2	4	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒ Operator Signature *H3*



### CONTROL SERUM

9	7	5	9	3	8	6	2	2	4	1	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒ Operator Signature *H3*



*This batch has been manufactured and checked according to quality requirements.  
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents. *H3*

DATE *12.06.2023*

### WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.