

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

QUALITY CONTROL CERTIFICATE

KIT NAME CARDIOLIPIN-M - CHORUS

Code 86048

Lot No. 943



2024-06



2023-06

REAGENTS

Label	Description	Lot.	Exp.	NO
D3331	CARDIOLIPIN-M	943	2024-06	6X6
B3331	CALIBRATORE CARDIOLIPIN-M	943	2024-06	1
C3331	CONTROL SERUM CARDIOLIPIN-M	943	2024-06	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	430	≥ 100
Control Serum (MPLU/mL)	52,0	27,0-73,3

CALIBRATOR


3	3	6	9	4	3	6	3	2	0	3	1	3	9	4	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature *HB*



CONTROL SERUM


3	3	6	9	4	3	6	1	8	3	4	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature *HB*



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents *[Signature]*

DATE 20-06-2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.