

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME MYCOPLASMA PNEUMONIAE IgM - CHORUS

Code 81035

Lot No. 928



2024-07



2023-05

REAGENTS

Label	Description	Lot.	Exp.	NO
D2261	MYCOPLASMA PNEUMONIAE IgM	928	2024-07	6x6
B2261	CALIBRATORE MYCOPLASMA PNEUMONIAE IgM	928	2024-07	1
C2261	CONTROL SERUM MYCOPLASMA PNEUMONIAE IgM	928	2024-07	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	590	≥ 100
Control Serum (INDEX)	1,5	1,1-2,0
Negative Control (INDEX)	0,2	$\leq 0,7$

CALIBRATOR

0	6	7	9	2	8	7	0	0	0	0	0	6	5	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---


Label checked with barcode reader YES ☒ Operator Signature *FB*



CONTROL SERUM

0	6	7	9	2	8	7	1	1	2	0	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒ Operator Signature *FB*



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents..... *Aluciel*

DATE 20/06/2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.