

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME CYTOMEGALOVIRUS IgG -CHORUS

Code	81010		
Lot No.	927	2024-04	2023-06

REAGENTS

Label	Description	Lot.	Exp.	NO
D2161	Cytomegalovirus IgG	927	2024-04	6x6
B2162	CALIBRATORE Cytomegalovirus IgG	927	2024-06	1
C2161	CONTROL SERUM Cytomegalovirus IgG	927	2024-06	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	320	≥ 100
Control Serum (IU/mL)	13,1	5,4-15,7
Negative Control (IU/mL)	<0,5	<0,5

CALIBRATOR

0	0	1	9	2	7	6	6	2	7	3	9	2	5	6	2	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒ Operator Signature



CONTROL SERUM

0	0	1	9	2	7	6	2	8	5	2	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒ Operator Signature



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents.....

DATE 28/06/2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.