

QUALITY CONTROL CERTIFICATE

KIT NAME 88000-G4 MEADOW FESCUE

Code 88000-G4

Lot No. 930

2024-04



2023-06

REAGENTS

Label	Description	Lot.	Exp.	NO
D0081	88000-G4 MEADOW FESCUE	930	2024-04	2X6
B0080	CALIBRATORE 88000-G4 MEADOW FESCUE	930	2024-06	1
C0080	CONTROL SERUM 88000-G4 MEADOW FESCUE	930	2024-06	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	279	> 100
Control Serum (KU/L)	17,5	10,1-44,7

CALIBRATOR

6	7	5	9	3	0	6	5	2	0	6	2	1	8	7	4	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---


Label checked with barcode reader

YES



Operator Signature

Operator Signature



CONTROL SERUM

6	7	5	9	3	0	6	3	9	7	2	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature

Operator Signature



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents.....

Signature

DATE 23.06.2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.