

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada del Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME CHORUS PROMONITOR ANTI-ADALIMUMAB

| | | | |
|----------------|-------|---------|---------|
| Code | 86703 | | |
| Lot No. | 934 | 2024-05 | 2023-05 |

REAGENTS

| Label | Description | Lot. | Exp | No |
|-------|----------------------------|------|---------|-----|
| D4340 | PROMONITOR ANTI-ADALIMUMAB | 934 | 2024-05 | 6X6 |
| B4340 | CALIBRATOR ANTI-ADALIMUMAB | 934 | 2025-02 | 1 |
| C4340 | CONTROL ANTI-ADALIMUMAB | 934 | 2025-02 | 1 |

CONTROLS

| Found results | Expected results |
|---------------|------------------|
|---------------|------------------|

| | | |
|------------------|------|-----------|
| Calibrator (mOD) | 360 | > 100 |
| Control (AU/ml) | 32,1 | 30,0-38,7 |


CALIBRATOR

| | | | | | | | | | | | | | | | | | |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|
| 2 | 7 | 7 | 9 | 3 | 4 | 6 | 5 | 1 | 4 | 4 | 5 | 6 | 0 | 9 | 7 | 0 | 0 |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|

Label checked with barcode reader

YES ☒

Operator Signature




CONTROL SERUM

| | | | | | | | | | | | | | | | | | |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|
| 2 | 7 | 7 | 9 | 3 | 4 | 6 | 1 | 4 | 2 | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 1 |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|

Label checked with barcode reader

YES ☒

Operator Signature



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents

DATE 30/06/2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.