

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME a-TPO - CHORUS

Code 86074/12

Lot No. 944

2024-06

2023-06

REAGENTS

Label	Description	Lot.	Exp.	NO
D3591	a-TPO	944	2024-06	2x6
B3591	CALIBRATORE a-TPO	944	2024-06	1
C3591	CONTROL SERUM a-TPO	944	2024-06	1

CONTROLS

	Found results	Expected results
Calibrator (mOD)	322	≥ 100
Control Serum (IU/mL)	411,3	233,3-533,3


CALIBRATOR

3 0 1 9 4 4 6 5 1 9 3 7 3 4 5 2 0 0

Label checked with barcode reader

YES

Operator Signature *FB*



CONTROL SERUM

3 0 1 9 4 4 6 2 1 3 8 0 0 0 0 0 0 1

Label checked with barcode reader

YES

Operator Signature *FB*



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents *FB*

DATE 03/07/2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.