



DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020

Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

QUALITY CONTROL CERTIFICATE

KIT NAME SARS-CoV-2 Antigen - CHORUS

Code 81410

Lot No. 926



2024-04



2023-05

REAGENTS

Label	Description	Lot.	Exp.	NO
D4220	SARS-CoV-2 Antigen	926	2024-04	6X6
B4220	CALIBRATORE SARS-CoV-2 Antigen	926	2024-05	1
C4220	CONTROL SARS-CoV-2 Antigen	926	2024-05	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	400	≥ 100
Control (AU/mL)	53,8	35,3-83,3

CALIBRATOR

1	4	0	9	2	6	6	4	2	0	3	7	3	0	4	9	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature MB



CONTROL SERUM


1	4	0	9	2	6	6	2	3	4	4	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature MB



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents.....

DATE 11/07/2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.