

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME ENA-6 S - CHORUS

Code 86012

Lot No. 936



2024-05



2023-05

REAGENTS

Label	Description	Lot.	Exp.	NO
D3131	ENA-6 S	936	2024-05	6X6
B3131	CALIBRATORE ENA-6 S	936	2024-05	1
C3131	CONTROL SERUM ENA-6 S	936	2024-05	1

CONTROLS

	Found results	Expected results
Calibrator (mOD)	440	≥ 100
Control Serum (ndx)	3,1	2,0-3,6

CALIBRATOR

1 0 2 9 3 6 6 0 0 0 0 0 3 6 0 0 0 0


Label checked with barcode reader

YES



Operator Signature

FB



CONTROL SERUM

1 0 2 9 3 6 6 2 0 3 6 0 0 0 0 0 0 1


Label checked with barcode reader

YES



Operator Signature

FB



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents.....

DATE 13/07/2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.