

DIESSE DIAGNOSTICA SENESE SpA  
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39  
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale  
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

## QUALITY CONTROL CERTIFICATE

KIT NAME BETA 2-GLYCOPROTEIN-G - CHORUS

Code 86050

Lot No. 944



2024-06



2023-06

### REAGENTS

Label	Description	Lot.	Exp.	NO
D3351	BETA 2-GLYCOPROTEIN-G	944	2024-06	6X6
B3351	CALIBRATORE BETA 2-GLYCOPROTEIN-G	944	2024-06	1
C3351	CONTROL SERUM BETA 2-GLYCOPROTEIN-G	944	2024-06	1

### CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	527	$\geq 100$
Control Serum (AU/mL)	28,4	23,8-47,8

### CALIBRATOR

2	0	6	9	4	4	6	2	2	6	3	5	5	3	5	1	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature *Green*



### CONTROL SERUM


2	0	6	9	4	4	6	1	4	2	5	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature *Green*



*This batch has been manufactured and checked according to quality requirements.  
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents.....

DATE

18/07/23

### WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.