

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020

DIESSE

Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME LEGIONELLA PNEUMOPHILA 1 IgG -CHORUS

Code 81091

Lot No. 952

2024-04

2023-07

REAGENTS

Label	Description	Lot.	Exp.	NO
D7371	LEGIONELLA PNEUMOPHILA 1 IgG	952	2024-04	6X6
B7371	CALIBRATORE L. PNEUMOPHILA 1 IgG	952	2024-04	1
C7371	CONTROL SERUM L. PNEUMOPHILA 1 IgG	952	2024-04	1

CONTROLS


	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	540	≥ 100
Control Serum (INDEX)	2,7	1,8-3,3
Negative Control (INDEX)	0,0	$< 0,7$

CALIBRATOR

1	5	5	9	5	2	5	0	0	0	0	0	3	9	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---


Label checked with barcode reader YES ☒ Operator Signature *HB*



CONTROL SERUM

1	5	5	9	5	2	5	1	8	3	3	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒ Operator Signature *HB*



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents *HB*

DATE 24/07/2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.