

DIESSE DIAGNOSTICA SENESE SpA  
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39  
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale  
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

## QUALITY CONTROL CERTIFICATE

KIT NAME RUBELLA IgG AVIDITY CHORUS

Code 81095

Lot No. 944



2024-06



2023-06

### REAGENTS

Label	Description	Lot.	Exp.	NO
D2501	RUBELLA IgG AVIDITY	944	2024-06	2x6
C2501	CONTROL SERUM HIGH RUBELLA IgG AVIDITY	944	2024-06	1
G2501	SAMPLE DILUENT	944	2024-09	1

### CONTROLS

	Found results	Expected results
--	---------------	------------------

Control Serum (%)	74,85%	> 45 ≤ 90 %
-------------------	--------	-------------

### CALIBRATOR

--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--

Label checked with barcode reader

YES

☐

Operator Signature .....

STICK  
LABEL

### CONTROL SERUM

0	8	5	9	4	4	6	4	5	9	0	0	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES

☒

Operator Signature .....



086944646900000001  
Rubella IgG Avidity

CONTROL HIGH 0.3 mL IV0

LOT B44

2024-06



2°C 8°C

Monteriggioni  
Italy

C2501

CE  
0123

*This batch has been manufactured and checked according to quality requirements.  
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents.....

DATE 24/07/2023

### WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.