

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME CHORUS PROMONITOR ANTI-VEDOLIZUMAB

Code 86711

Lot No. 943



2024-06



2023-06

REAGENTS

Label	Description	Lot.	Exp.	NO
D4490	PROMONITOR ANTI-VEDOLIZUMAB	943	2024-06	6x6
B4490	CALIBRATORE ANTI-VEDOLIZUMAB	943	2025-03	1
C4490	CONTROL SERUM ANTI-VEDOLIZUMAB	943	2025-03	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	243	≥ 100
Control Serum (AU/mL)	66,0	57,6-87,7

CALIBRATOR

2 7 9 9 4 3 6 5 1 9 3 3 3 7 5 9 0 0


Label checked with barcode reader

YES



Operator Signature

AB



CONTROL SERUM

2 7 9 9 4 3 6 2 2 3 3 0 0 0 0 0 0 1


Label checked with barcode reader

YES



Operator Signature

AB



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents

DATE

24.07.2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.