

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME ANA-Screen - CHORUS

Code 86014

Lot No. 946



2024-06



2023-06

REAGENTS

Label	Description	Lot.	Exp.	NO
D3141	ANA-Screen	946	2024-06	6X6
B3141	CALIBRATORE ANA-Screen	946	2024-06	1
C3141	CONTROL SERUM ANA-Screen	946	2024-06	1

CONTROLS

	Found results	Expected results
Calibrator (mOD)	420	≥ 100
Control Serum (Ndx)	3,5	2,3-4,5


CALIBRATOR

1	1	5	9	4	6	6	0	0	0	0	0	3	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES ☒

Operator Signature




CONTROL SERUM

1	1	5	9	4	6	6	2	3	4	3	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES ☒

Operator Signature



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents.....

DATE 20/07/2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.