

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME TG Extended Range

Code	86110		
Lot No.	943	2024-06	2023-07

REAGENTS

Label	Description	Lot.	Exp.	NO
D4050	TG Extended Range	943	2024-06	6x6
B4050	CALIBRATOR TG Extended Range	943	2024-07	1
C4050	CONTROL SERUM TG Extended Range	943	2024-07	1


CONTROLS

	Found results	Expected results
CALIBRATOR (mOD)	444	≥ 100
CONTROL SERUM (ng/ml)	2,6	1,5-3,9

CALIBRATOR

4	1	8	9	4	3	6	7	1	3	2	1	3	9	4	8	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---


Label checked with barcode reader YES ☒ Operator Signature



CONTROL SERUM

4	1	8	9	4	3	6	1	8	3	4	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒ Operator Signature



This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents

DATE 24/07/2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.