

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME **a-TPO - CHORUS**

Code **86074**

Lot No. **945**

2024-06

2023-06

REAGENTS

Label	Description	Lot.	Exp.	NO
D3591	a-TPO	945	2024-06	6X6
B3591	CALIBRATORE a-TPO	945	2024-06	1
C3591	CONTROL SERUM a-TPO	945	2024-06	1

CONTROLS

	Found results	Expected results
Calibrator (mOD)	360	≥ 100
Control Serum (IU/mL)	436,1	233,3-533,3

CALIBRATOR

3	0	1	9	4	5	6	5	1	9	3	7	3	4	5	2	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature



CONTROL SERUM


3	0	1	9	4	5	6	2	1	3	8	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents

DATE **26.07.2023**

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.