

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI) – Italia – Strada dei Laghi 39
Tel +39-0577-319554 Fax +39-0577-319020



Pag 1/1
Sede Legale e Amministrazione
20144 Milano (Italy) Via A. Solari 19 sc. 6
Tel. +39-02-4859121 r.a. Fax +39-02-48008530

QUALITY CONTROL CERTIFICATE

KIT NAME: 88000- C2- PENICILLOYL V

Code 88000 - C2

Lot No. 943



2024-06



2023-07

REAGENTS

Label	Description	Lot.	Exp	No
D4931	88000- C2 PENICILLOYL V	943	2024-06	2X6
B4930	CALIBRATORE 88000- C2 PENICILLOYL V	943	2024-06	1
C4930	CONTROL SERUM 88000- C2 PENICILLOYL V	943	2024-06	1

CONTROLS

Found results	Expected results
---------------	------------------

Calibrator (mOD)	290	≥ 100
Control Serum (KU/L)	17,2	10,0-41,5

CALIBRATOR

6	6	2	9	4	3	6	5	2	1	6	6	1	7	8	1	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---


Label checked with barcode reader

YES



Operator Signature

RB



CONTROL SERUM

6	6	2	9	4	3	6	4	1	7	7	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---


Label checked with barcode reader

YES



Operator Signature

RB



The test has been carried out in accordance with the package insert

TESTED BY: Simone Balocchi

Simone Balocchi

APPROVED

RESPONSIBLE HEAD, Quality Control Reagents

Blue

DATE

28.07.2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested by an FDA approved method and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.