

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME MYCOPLASMA IgA - CHORUS

Code 81033

Lot No. 928

2024-04



2023-04

REAGENTS

Label	Description	Lot.	Exp.	NO
D7551	MYCOPLASMA IgA	928	2024-04	6X6
B7551	CALIBRATORE MYCOPLASMA IgA	928	2024-04	1
C7551	CONTROL SERUM MYCOPLASMA IgA	928	2024-04	1

CONTROLS


	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	330	≥ 100
Control Serum (AU/mL)	44,9	28,6-48,7
Negative Control (AU/mL)	<10	<10

CALIBRATOR

0 6 8 9 2 8 6 8 1 2 1 9 4 8 7 4 0 0


Label checked with barcode reader YES ☒ Operator Signature HB



CONTROL SERUM

0 6 8 9 2 8 6 1 5 2 7 0 0 0 0 0 0 1

Label checked with barcode reader YES ☒ Operator Signature HB



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents. Alfano

DATE 28/07/2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.