

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME 88000-F8 MAIZE

Code 88000- F8

Lot No. 940



2024-05



2023-07

REAGENTS

Label	Description	Lot.	Exp.	NO
D0511	88000-F8 MAIZE	940	2024-05	2x6
B0510	CALIBRATORE 88000-F8 MAIZE	940	2024-06	1
C0510	CONTROL SERUM 88000-F8 MAIZE	940	2024-06	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	290	≥ 100
Control Serum (KU/L)	17,2	10,0-41,5

CALIBRATOR

6	6	6	9	4	0	6	5	2	1	6	6	1	7	8	1	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature



CONTROL SERUM


6	6	6	9	4	0	6	4	1	7	7	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature



This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents

DATE 28/07/2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.