

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME RUBELLA IgM - CHORUS

Code 81031

Lot No. 947



2024-06



2023-07

REAGENTS

Label	Description	Lot.	Exp.	NO
D2241	RUBELLA IgM	947	2024-06	6X6
B2241	CALIBRATORE RUBELLA IgM	947	2024-07	1
C2241	CONTROL SERUM	947	2024-07	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	630	≥ 100
Control Serum (INDEX)	2,3	1,6-3,0
Negative Control (INDEX)	0,0	$< 0,7$

CALIBRATOR


1	3	5	9	4	7	6	0	0	0	0	0	4	4	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature *SB*



CONTROL SERUM

1	3	5	9	4	7	6	1	6	3	0	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature *SB*



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents..... *Aliaudi*

DATE *07-08-2023*

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent,