

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME EPSTEIN-BARR EARLY ANTIGEN IgM

Code 81059

Lot No. 944



2024-06



2023-06

REAGENTS

Label	Description	Lot.	Exp.	NO
D2382	Epstein-Barr Early Antigen IgM	944	2024-06	6x6
B2382	CALIBRATORE Epstein-Barr Early Antigen IgM	944	2024-06	1
C2382	CONTROL SERUM Epstein-Barr Early Antigen IgM	944	2024-06	1

CONTROLS

	Found results	Expected results
Calibrator (mOD)	320	≥ 100
Control Serum (Index)	3,3	1,9-3,5
Negative Control (Index)	0,0	$< 0,7$

CALIBRATOR

0	2	9	9	4	4	6	0	0	0	0	0	3	7	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature MB



CONTROL SERUM


0	2	9	9	4	4	6	1	9	3	5	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature MB



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents..... Aliaudi

DATE 28/08/2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive