

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME SS-A - CHORUS

Code	86018
Lot No.	928
	2024-07
	2023-04

REAGENTS

Label	Description	Lot.	Exp.	NO
D3181	SS-A	928	2024-07	6X6
B3181	CALIBRATORE SS-A	928	2024-07	1
C3181	CONTROL SERUM SS-A	928	2024-07	1


CONTROLS

	Found results	Expected results
Calibrator (mOD)	413	≥ 100
Control Serum (AU/mL)	54,6	36,0-71,7

CALIBRATOR

1	9	3	9	2	8	7	1	1	6	2	6	4	0	5	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒ Operator Signature



CONTROL SERUM

1	9	3	9	2	8	7	1	8	3	3	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒ Operator Signature



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents

DATE 23/08/2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.