

QUALITY CONTROL CERTIFICATE

KIT NAME CHLAMYDOPHILA PNEUMONIAE IgM - CHORUS

Code 81251

Lot No. 945



2024-06



2023-06

REAGENTS

Label	Description	Lot.	Exp.	NO
D8600	Chlamydomphila pneumoniae IgM	945	2024-06	6X6
B8600	CALIBRATORE Chlamydomphila p. IgM	945	2024-06	1
C8600	CONTROL SERUM Chlamydomphila p. IgM	945	2024-06	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	410	≥ 100
Control Serum (Index)	3,2	2,2-4,1
Negative Control (Index)	0,1	$\leq 0,7$

CALIBRATOR


2	3	0	9	4	5	6	0	0	0	0	0	3	2	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature *Open*



CONTROL SERUM


2	3	0	9	4	5	6	2	2	4	1	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature *Open*



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents *Open*

DATE 22-08-2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.