

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME CHLAMYDOPHILA PNEUMONIAE IgA - CHORUS

Code 81252

Lot No. 953



2024-07



2023-07

REAGENTS

Label	Description	Lot.	Exp.	NO
D8610	Chlamydomphila pneumoniae IgA	953	2024-07	6X6
B8610	CALIBRATORE Chlamydomphila p. IgA	953	2024-07	1
C8610	CONTROL SERUM Chlamydomphila p. IgA	953	2024-07	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	410	≥ 100
Control Serum (AU/mL)	30,1	15,0-37,5
Negative Control (AU/mL)	$<10,0$	$\leq 10,0$

CALIBRATOR

1 9 9 9 5 3 6 8 1 1 1 9 7 1 2 9 0 0

Label checked with barcode reader

YES



Operator Signature *FB*



CONTROL SERUM

1 9 9 9 5 3 6 1 0 1 8 0 0 0 0 0 0 1

Label checked with barcode reader

YES



Operator Signature *FB*



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents..... *Alciandri*

DATE *28/08/2023*

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.