

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME Scl-70 - CHORUS

Code 86026

Lot No. 954



2024-07



2023-07

REAGENTS

Label	Description	Lot.	Exp.	NO
D3221	Scl-70	954	2024-07	6X6
B3221	CALIBRATORE Scl-70	954	2024-07	1
C3221	CONTROL SERUM Scl-70	954	2024-07	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	320	≥ 100
Control Serum (AU/mL)	73,0	40,9-100,0

CALIBRATOR

1	9	5	9	5	4	6	2	1	9	3	0	3	0	4	3	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature *HB*



CONTROL SERUM


1	9	5	9	5	4	6	2	3	4	3	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature *HB*



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents *HB*

DATE 30/08/2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive.