

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME SS-B - CHORUS

Code 86024

Lot No. 945



2024-06



2023-06

REAGENTS

| Label | Description | Lot. | Exp. | NO |
|-------|--------------------|------|---------|-----|
| D3201 | SS-B | 945 | 2024-06 | 6X6 |
| B3201 | CALIBRATORE SS-B | 945 | 2024-06 | 1 |
| C3201 | CONTROL SERUM SS-B | 945 | 2024-06 | 1 |

CONTROLS

| | Found results | Expected results |
|--|---------------|------------------|
|--|---------------|------------------|

| | | |
|-----------------------|------|------------|
| Calibrator (mOD) | 460 | ≥ 100 |
| Control Serum (AU/mL) | 37,8 | 25,5-52,5 |

CALIBRATOR

| | | | | | | | | | | | | | | | | | |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|
| 1 | 9 | 4 | 9 | 4 | 5 | 6 | 3 | 2 | 0 | 3 | 6 | 4 | 1 | 5 | 8 | 0 | 0 |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|

Label checked with barcode reader

YES



Operator Signature

FB



CONTROL SERUM

| | | | | | | | | | | | | | | | | | |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|
| 1 | 9 | 4 | 9 | 4 | 5 | 6 | 1 | 7 | 3 | 2 | 0 | 0 | 0 | 0 | 0 | 0 | 1 |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|

Label checked with barcode reader

YES



Operator Signature

FB



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents

FB

DATE 31/08/2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.