

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME RUBELLA IgG - CHORUS

Code 81030

Lot No. 953

2024-07

2023-07

REAGENTS

Label	Description	Lot.	Exp.	NO
D2231	RUBELLA IgG	953	2024-07	6X6
B2232	CALIBRATORE RUBELLA IgG	953	2024-07	1
C2231	CONTROL SERUM RUBELLA IgG	953	2024-07	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	538	≥ 100
Control Serum (IU/mL)	22,5	18,0-37,5
Negative Control (IU/mL)	$<5,0$	< 5


CALIBRATOR

0 8 1 9 5 3 6 6 2 5 4 1 4 0 4 7 0 0

Label checked with barcode reader

YES ☒

Operator Signature



CONTROL SERUM

0 8 1 9 5 3 6 1 8 3 3 0 0 0 0 0 0 1

Label checked with barcode reader

YES ☒

Operator Signature



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents

DATE 31/08/2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.