

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020

DIESSE

Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME TREPONEMA IgM -CHORUS

Code 81051

Lot No. 945

2024-09

2023-07

REAGENTS

Label	Description	Lot.	Exp.	NO
D2331	TREPONEMA IgM	945	2024-09	6X6
B2331	CALIBRATORE TREPONEMA IgM	945	2024-10	1
C2331	CONTROL SERUM TREPONEMA IgM	945	2024-10	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	360	≥ 100
Control Serum (INDEX)	2,2	1,7-3,1
Negative Control (INDEX)	0,4	$< 0,7$

CALIBRATOR

0 1 5 9 4 5 7 0 0 0 0 0 4 2 0 0 0 0 0

Label checked with barcode reader

YES



Operator Signature

[Signature]



CONTROL SERUM

0 1 5 9 4 5 7 1 7 3 1 0 0 0 0 0 0 0 1


Label checked with barcode reader

YES



Operator Signature

[Signature]



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents

[Signature]

DATE 31/08/2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.