

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020

DIESSE

Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME Epstein-Barr VCA IgM II - CHORUS

Code 81054

Lot No. 945

2024-09

2023-06

REAGENTS

Label	Description	Lot	Exp.	NO
D9490	Epstein-Barr VCA IgM II	945	2024-09	6x6
B9490	CALIBRATOR Epstein-Barr VCA IgM II	945	2024-09	1
C9490	CONTROL SERUM Epstein-Barr VCA IgM II	945	2024-09	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	484	≥ 100
Control Serum (INDEX)	2,4	1,8-3,3
Negative Control (INDEX)	0,1	$< 0,7$

CALIBRATOR

0 3 0 9 4 5 7 0 0 0 0 0 4 0 0 0 0 0 0

Label checked with barcode reader

YES ☒

Operator Signature *88*



CONTROL SERUM

0 3 0 9 4 5 7 1 8 3 3 0 0 0 0 0 0 0 1

Label checked with barcode reader

YES ☒

Operator Signature *88*



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents *88*

DATE *20.09.23*

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.