

DIESSE DIAGNOSTICA SENESE SpA  
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39  
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale  
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

## QUALITY CONTROL CERTIFICATE

KIT NAME VARICELLA IgG - CHORUS

Code 81078

Lot No. 948

2024-06

2023-06

### REAGENTS

| Label | Description                 | Lot. | Exp.    | NO  |
|-------|-----------------------------|------|---------|-----|
| D2471 | VARICELLA IgG               | 948  | 2024-06 | 6x6 |
| B2472 | CALIBRATORE VARICELLA IgG   | 948  | 2024-06 | 1   |
| C2471 | CONTROL SERUM VARICELLA IgG | 948  | 2024-06 | 1   |

### CONTROLS


|  | Found results | Expected results |
|--|---------------|------------------|
|--|---------------|------------------|

|                          |     |            |
|--------------------------|-----|------------|
| Calibrator (mOD)         | 410 | $\geq 100$ |
| Control Serum (Index)    | 3,6 | 2,4-4,5    |
| Negative Control (Index) | 0,0 | $< 0,7$    |

### CALIBRATOR

|   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|
| 0 | 1 | 4 | 9 | 4 | 8 | 6 | 0 | 0 | 0 | 0 | 0 | 2 | 9 | 0 | 0 | 0 | 0 |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|


Label checked with barcode reader YES ☒ Operator Signature *SB*



### CONTROL SERUM

|   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|
| 0 | 1 | 4 | 9 | 4 | 8 | 6 | 2 | 4 | 4 | 5 | 0 | 0 | 0 | 0 | 0 | 0 | 1 |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|

Label checked with barcode reader YES ☒ Operator Signature *SB*



This batch has been manufactured and checked according to quality requirements.  
Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents *SB*

DATE 05.09.2023

### WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.