

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME CHORUS ECHINOCOCCUS IgG

Code	81284
Lot No.	961
	2024-08
	2023-08

REAGENTS

Label	Description	Lot.	Exp.	NO
D0010	Echinooccus IgG	961	2024-08	2X6
B0010	CALIBRATOR Echinooccus IgG	961	2024-08	1
C0010	CONTROL SERUM Echinooccus IgG	961	2024-08	1

CONTROLS


	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	600	≥ 100
Control Serum (Index)	1,4	1,0-1,8

CALIBRATOR

0	3	5	9	6	1	6	0	0	0	0	0	7	2	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---


Label checked with barcode reader YES ☒ Operator Signature



CONTROL SERUM

0	3	5	9	6	1	6	1	0	1	8	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒ Operator Signature



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents

DATE 05/09/2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.