

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME TICK-BORNE ENCEPHALITIS VIRUS IgG - CHORUS

Code 81276

Lot No. 963

2024-08

2023-09

REAGENTS

Label	Description	Lot.	Exp.	NO
D0140	TICK-BORNE ENCEPHALITIS VIRUS IgG	963	2024-08	6x6
B0141	CALIBRATORE TICK-BORNE ENCEPHALITIS VIRUS IgG	963	2024-12	1
C0141	CONTROL SERUM TICK-BORNE ENCEPHALITIS VIRUS IgG	963	2024-12	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	460	≥ 100
Control (AU/mL)	62,4	46,4-72,0


CALIBRATOR

9	6	6	9	6	3	6	1	1	6	4	1	3	6	7	4	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES ☒

Operator Signature *88*




CONTROL SERUM

9	6	6	9	6	3	6	2	0	3	6	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES ☒

Operator Signature *88*



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents *88*

DATE *06.09.23*

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.