

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020

DIESSE

Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME AMA-M2 - CHORUS

Code 86084

Lot No. 952

2024-10

2023-09

REAGENTS

Label	Description	Lot.	Exp.	NO
D3651	AMA-M2	952	2024-10	6X6
B3651	CALIBRATORE AMA-M2	952	2024-11	1
C3651	CONTROL SERUM AMA-M2	952	2024-11	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	430	≥ 100
Control Serum (AU/mL)	70,1	39,1-121,1

CALIBRATOR

7 0 5 9 5 2 7 2 6 3 8 3 3 5 7 0 0

Label checked with barcode reader

YES ☒

Operator Signature *88*




CONTROL SERUM

7 0 5 9 5 2 7 2 1 4 0 0 0 0 0 0 0 1

Label checked with barcode reader

YES ☒

Operator Signature *88*



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents *88*

DATE *16.09.2023*

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.