

DIESE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

PRODUCT NAME: VACU-CODE

Code	10230		
Lot No.	3723	2024-08	2023-09
Reference IO-09-88			

ON LINE QUALITY CONTROLS (model CDP not enclosed)

QC Passed

- sodium citrate solution concentration ☒
- sodium citrate volume ☒
- vacuum level (*) ☒
- correct placement of the cap ☒
- no extra body inside the product ☒
- (*) Only for VACU-CODE and VACU-TEC tubes ☒

PACKAGING QUALITY CONTROL (model CDP not enclosed)

QC Passed

- control of the number of the product for each pack ☒

QUALITY CONTROLS ON FINISHED PRODUCT (model CDP not enclosed)

QC Passed

- pack integrity ☒
- data printed on the product label ☒

QC Passed

N/A

Irradiation treatment Certificate n° date



QC Passed

N/A

Gamma rays treatment process in compliance with the specifications of the supply Agreement



Document number:

ODV N.A.

This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.

APPROVED

RESPONSIBLE HEAD, Quality Control Reagents.

DATE 15/09/23