

DIESSE DIAGNOSTICA SENESE SpA  
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39  
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale  
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

## QUALITY CONTROL CERTIFICATE

KIT NAME DEAMIDATED GLIADIN-A - CHORUS

Code	86098		
Lot No.	961	2024-08	2023-08

### REAGENTS

Label	Description	Lot.	Exp.	NO
D8650	Deamidated Gliadin-A	961	2024-08	6X6
B8650	CALIBRATORE Deamidated Gliadin-A	961	2024-08	1
C8650	CONTROL SERUM Deamidated Gliadin-A	961	2024-08	1

### CONTROLS


	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	390	$\geq 100$
Control Serum (AU/mL)	166,7	79,2-400,0

### CALIBRATOR

2	5	4	9	6	1	6	2	2	4	3	6	2	3	5	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---


Label checked with barcode reader YES ☒ Operator Signature



### CONTROL SERUM

2	5	4	9	6	1	6	3	1	5	7	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒ Operator Signature



*This batch has been manufactured and checked according to quality requirements.  
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents

DATE 19/09/2023

### WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.