

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME RESPIRATORY SYNCYTIAL VIRUS IgA - CHORUS

Code 81037

Lot No. 970

2024-06

2023-09

REAGENTS

Label	Description	Lot.	Exp.	NO
D7671	Respiratory Syncytial Virus IgA	970	2024-06	6X6
B7671	CALIBRATORE Respiratory Syncytial Virus IgA	970	2024-06	1
C7671	CONTROL SERUM Respiratory Syncytial Virus IgA	970	2024-06	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	350	≥ 100
Control Serum (Ndx)	2,2	1,6-3,0
Negative Control (Ndx)	0,1	$\leq 0,7$

CALIBRATOR

0	7	8	9	7	0	5	0	0	0	0	0	4	3	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES ☒

Operator Signature




CONTROL SERUM

0	7	8	9	7	0	5	1	6	3	0	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES ☒

Operator Signature



This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents.....

DATE 25/09/2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.