

QUALITY CONTROL CERTIFICATE

KIT NAME MEASLES IgG CHORUS

Code 81072

Lot No. 970



2024-12



2023-09

REAGENTS

Label	Description	Lot.	Exp.	NO
D2431	MEASLES IgG	970	2024-12	6X6
B2432	CALIBRATORE MEASLES IgG	970	2024-12	1
C2431	CONTROL SERUM MEASLES IgG	970	2024-12	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	340	≥ 100
Control Serum (Index)	4,0	2,6-4,9
Negative Control (Index)	0,1	$< 0,7$

CALIBRATOR

0	0	9	9	7	0	7	0	0	0	0	0	2	7	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature

Oster



CONTROL SERUM

0	0	9	9	7	0	7	2	6	4	9	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---


Label checked with barcode reader

YES



Operator Signature

Oster



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents

Oster

DATE 26.09.2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.