

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020

DIESSE

Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME HSV 2 SCREEN - CHORUS

Code 81019

Lot No. 944

2024-09

2023-07

REAGENTS

Label	Description	Lot.	Exp.	NO
D2201	HSV 2 SCREEN			
B2201	CALIBRATORE HSV 2 SCREEN	944	2024-09	6x6
C2201	CONTROL SERUM HSV 2 SCREEN	944	2024-09	1
		944	2025-06	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	520	≥ 100
Control Serum (index)	1,5	1,3-8,0
Negative Control (index)	0,7	< 0,7

CALIBRATOR

0	3	4	9	4	4	7	0	0	0	0	0	6	9	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature *HB*



CONTROL SERUM

0	3	4	9	4	4	7	1	3	8	0	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature *HB*



**This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.**

APPROVED RESPONSIBLE HEAD, Quality Control Reagents. *HB*

DATE 21/05/2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.