

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020

DIESSE

Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME 88000-C204 - AMOXICILLIN-HSA

Code	88000-C204
Lot No.	963
	 2024-08
	 2023-09

REAGENTS

Label	Description	Lot.	Exp	No
D3741	88000- C204 AMOXICILLIN-HSA	963	2024-08	2X6
B3740	CALIBRATORE 88000- C204 AMOXICILLIN-HSA	963	2024-10	1
C3740	CONTROL SERUM 88000-C204 AMOXICILLIN-HSA	963	2024-10	1

CONTROLS

Found results	Expected results
---------------	------------------

Calibrator (mOD)	242	> 100
Control Serum (KU/L)	12,1	7,5-15,7

CALIBRATOR


6	9	7	9	6	3	6	5	1	8	6	2	2	4	8	7	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature



CONTROL SERUM


6	9	7	9	6	3	6	3	0	5	5	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution as Research Use Only (RUO) product.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents

DATE 28/09/2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.