

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020

DIESSE

Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME **tTg-A - CHORUS**

Code **86058**

Lot No. **955-A**

2024-07

2023-07

REAGENTS

Label	Description	Lot.	Exp.	NO
D3431	tTg-A			
B3431	CALIBRATORE tTg-A	955	2024-07	6X6
C3431	CONTROL SERUM tTg-A	955-1	2024-07	1
		955-1	2024-07	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	469	≥ 100
Control Serum (AU/mL)	27,5	20,8-41,5

CALIBRATOR

3	3	9	9	5	5	6	4	2	3	3	6	4	7	4	3	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES ☒

Operator Signature




CONTROL SERUM

3	3	9	9	5	5	6	1	5	2	8	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES ☒

Operator Signature



**This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.**

APPROVED RESPONSIBLE HEAD, Quality Control Reagents.

DATE **29.09.2023**

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.