

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020

DIESSE

Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME PR3 - CHORUS

Code	86066		
Lot No.	970	2024-09	2023-09

REAGENTS

Label	Description	Lot.	Exp.	NO
D3512	PR3	970	2024-09	6X6
B3512	CALIBRATORE PR3	970	2024-09	1
C3512	CONTROL SERUM PR3	970	2024-09	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	370	≥ 100
Control Serum (AU/mL)	78,5	56,0-100,0

CALIBRATOR

2	5	2	9	7	0	6	4	1	4	2	9	2	7	4	9	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒ Operator Signature

252970641429274900
PR3
CALIBRATOR 0.175 mL [IVD]
LOT 970
 2024-09 2°C - 8°C
DIESSE Monteriggioni Italy CE

CONTROL SERUM

2	5	2	9	7	0	6	2	7	4	9	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒ Operator Signature

252970627490000001
PR3
CONTROL 0.425 mL [IVD]
LOT 970
 2024-09 2°C - 8°C
DIESSE Monteriggioni Italy CE

*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents

DATE 02/10/2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.