

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME snRNP-C - CHORUS

Code	86092		
Lot No.	970	2024-09	2023-09

REAGENTS

Label	Description	Lot.	Exp.	NO
D7441	snRNP-C	970	2024-09	6X6
B7441	CALIBRATORE snRNP-C	970	2024-09	1
C7441	CONTROL SERUM snRNP-C	970	2024-09	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	350	≥ 100
Control Serum (AU/mL)	75,8	42,0-88,6

CALIBRATOR

1	9	1	9	7	0	6	1	1	5	2	5	3	7	3	9	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒ Operator Signature *MB*



CONTROL SERUM

1	9	1	9	7	0	6	1	9	3	5	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒ Operator Signature *MB*



This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents *MB*

DATE 02/10/2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive.