

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME DEAMIDATED GLIADIN-G - CHORUS

Code 86100

Lot No. 973

2024-09

2023-09

REAGENTS

Label	Description	Lot.	Exp.	NO
D8660	Deamidated Gliadin-G	973	2024-09	6X6
B8660	CALIBRATORE Deamidated Gliadin-G	973	2024-09	1
C8660	CONTROL SERUM Deamidated Gliadin-G	973	2024-09	1

CONTROLS

	Found results	Expected results
--	---------------	------------------


Calibrator (mOD)	590	≥ 100
Control Serum (AU/mL)	107,3	55,6-194,7

CALIBRATOR

2	5	5	9	7	3	6	1	1	8	2	7	3	6	4	6	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES ☒

Operator Signature 




CONTROL SERUM

2	5	5	9	7	3	6	1	9	3	6	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES ☒

Operator Signature 



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents 

DATE 11.10.2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.