

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME Vitamin D Total- CHORUS

Code 86900

Lot No. 953

2024-07

2023-09

REAGENTS

Label	Description	Lot.	Exp.	NO
D9702	25OH Vitamin D Total	953	2024-07	6x6
B9703	CALIBRATOR 25OH Vitamin D Total	953	2025-05	1
C9703	CONTROL SERUM 25OH Vitamin D Total	953	2025-05	4

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	383	≥ 100
CONTROL SERUM (ng/ml)	34,6	29,2-37,8

CALIBRATOR

8	3	5	9	5	3	6	3	3	5	4	0	5	3	6	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---


Label checked with barcode reader

YES



Operator Signature

88



CONTROL SERUM

8	3	5	9	5	3	6	1	3	2	5	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---


Label checked with barcode reader

YES



Operator Signature

88



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents

Bene

DATE 17.10.2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.