

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME MEASLES IgM CHORUS

Code 81073

Lot No. 944



2024-09



2023-06

REAGENTS

Label	Description	Lot.	Exp.	NO
D2442	MEASLES IgM	944	2024-09	6x6
B2442	CALIBRATORE MEASLES IgM	944	2024-09	2
C2442	CONTROL SERUM MEASLES IgM	944	2024-09	2

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	559	≥ 100
Control Serum (Index)	2,1	1,7-3,1
Negative Control (Index)	0,0	$< 0,7$

CALIBRATOR


1	3	3	9	4	4	7	0	0	0	0	0	4	2	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature *sf*



CONTROL SERUM

1	3	3	9	4	4	7	1	7	3	1	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature *sf*



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents. *sf*

DATE 18/10/2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.